181020110212

中国认可
国际互认
检测
TESTING
CNAS L0841

江苏省计量科学研究院

JIANGSU INSTITUTE OF METROLOGY

检 测 报 告

Test Report

报 告 编 号：T2020-40184
Report No.

委 托 单 位 Customer	江苏安科瑞电器制造有限公司 Jiangsu Acrel Electrical Manufacturing. Co., Ltd.
委托单位地址 Customer Add.	江苏省江阴市南闸街道东盟路 5 号 No.5, Dongmeng Road, Nanzha Street, Jiangyin City, Jiangsu Province
样 品 名 称 Name of Instrument	三相四线电子式电能表（导轨式） Three Phase Four Wire Static Energy Meter(DIN-rail mounted）
制 造 单 位 Manufacturer	江苏安科瑞电器制造有限公司 Jiangsu Acrel Electrical Manufacturing. Co., Ltd.

地址：南京市栖霞区文澜路 95 号（总部） 电子信箱：jsimguest@jsim.com.cn 网址：www.jsim.com.cn
Add: No.95, WenLan Road, XianLin University City,Nanjing（Headquarter） E-mail Website
业务电话：(025) 84636996 监督电话：(025)86435803 传真：(025)86435543 邮编：210023
Vocational Tel Supervising Tel Fax Post Code

LRI Distribuída no Brasil pela LRI | vendas@lri.com.br | 51 2129-3000

样品信息
Description of samples

名称 Name	三相四线电子式电能表(导轨式) Three Phase Four Wire Static Energy Meter(DIN-rail mounted)	型号/规格 Type/specification	ADW300 3×220/380V 3×1(6)A
编号 Serial No.	71912032800061	测量范围 Measuring Range	-
委托日期 Accepting date	2020-03-25	准确度等级/最大允许误差 Class of Accuracy/Maximum Permissible Errors	有功电能 0.5S 级 Active energy class 0.5S
到样日期 Date of samples received	2020-03-25	受样方式 Accepting way of samples	来样 Incoming sample
备注 Remarks	此类架装式仪表未进行防尘防水试验，使用时应安装于不低于 IP51 或 IP54 防护等级的封闭外罩内。 This kind of rack mounted meters were not done the dustproof and waterproof test.The meters should be installed in a closed enclosure not less than IP51 or IP54 protection level when they used.		

本次检测的依据：
Reference documents for the test

序号	文件代号及名称
1	IEC 62052-11:2003 *Electricity metering equipment (a.c.) – General requirements, tests and test conditions - Part 11: Metering equipment*
2	IEC 62053-22:2003 *Electricity metering equipment (a.c.) – Particular requirements – Part 22:Static meters for active energy(classes 0.2S and 0.5S)*

检测结论：
Results of test

依据用户委托，对样品进行了检测，所检项目符合 IEC 62052-11:2003 *Electricity metering equipment (a.c.) – General requirements, tests and test conditions - Part 11: Metering equipment* 和 IEC 62053-22:2003 *Electricity metering equipment (a.c.) – Particular requirements – Part 22:Static meters for active energy(classes 0.2S and 0.5S)*的要求。

We tested the samples according to the items of test standard. The test results meet the requirements of the following documents: IEC 62052-11:2003 *Electricity metering equipment(a.c.) – General requirements, tests and test conditions - Part 11: Metering equipment* and IEC 62053-22:2003 *Electricity metering equipment(a.c.) – Particular requirements – Part 22:Static meters for active energy(classes 0.2S and 0.5S).*

注：本报告提供的结果仅对本次被测样品有效。
The results in the report are only valid for the tested sample(s).

批准人 Approved by 邓凌翔

审核人 Checked by 李 博

检测人 Tested by 韩 沁

签发日期 Date of Issue 2020 年 Year 06 月 Month 30 日 Day

（检测专用章） Stamp

LRI Distribuída no Brasil pela LRI | vendas@lri.com.br | 51 2129-3000